

Fluorescence Spectrophotometer

Jasco FP6500

Validation and Correction

30th March, 2014

Ankit Raj

1. Validation


(by inbuilt software and given accessories)

Includes

- Wavelength accuracy
- Wavelength repeatability
- Resolution
- Stray light
- Minimum detectable amount
- Photometric stability

1.1 Wavelength accuracy

- Various wavelength light is sent via diffusion plate, and recorded as emission.



1.1 Wavelength accuracy

INSPECTION SHEET

Date: 30 / Mar / 2014

Model: FP-6500
Serial No.: C089160822
Room temp.: 24.9 C
Humidity: 41 %
Operator: Ankit

Inspected by:



Wavelength accuracy
Acceptance criterion





Overall Pass/Fail: Pass

(Ex) 253.7 +/-1.5 nm, 365.0 +/-1.5 nm, 435.8 +/-1.5 nm, 546.1 +/-1.5 nm

(Em) 253.7 +/-1.5 nm, 365.0 +/-1.5 nm, 435.8 +/-1.5 nm, 546.1 +/-1.5 nm


Result(s)

Standard value	Measured value	Difference
468.5nm(Ex)	467.6 nm	-0.9 nm
253.7nm(Ex)	253.7 nm	-0.0 nm
365.0nm(Ex)	364.6 nm	-0.4 nm
435.8nm(Ex)	435.6 nm	-0.2 nm
546.1nm(Ex)	546.1 nm	-0.0 nm
468.5nm(Em)	468.2 nm	-0.3 nm
253.7nm(Em)	253.9 nm	0.2 nm
365.0nm(Em)	365.1 nm	0.1 nm
435.8nm(Em)	435.8 nm	-0.0 nm
546.1nm(Em)	546.3 nm	0.2 nm



1.2 Wavelength repeatability

- 546.1 nm light is sent via diffusion plate, and recorded as emission. The repeatability of the light at 546.1nm is tested.



1.2 Wavelength repeatability

INSPECTION SHEET

Date: 30 / Mar / 2014

Model: FP-6500
Serial No.: C089160822
Room temp.: 24.9 C
Humidity: 41 %
Operator: Ankit

Inspected by:



Wavelength repeatability

Overall Pass/Fail: Pass
(Ex) +/-0.3 nm (Em) +/-0.3 nm



Mid point.
546.1 nm

Acceptance criterion
Result(s)

(Ex) Measured value
Difference
(Em) Measured value
Difference


546.1 nm	546.1 nm	546.1 nm
0.0 nm	0.0 nm	0.0 nm
546.3 nm	546.3 nm	546.3 nm
0.0 nm	0.0 nm	0.0 nm

546.3 nm



1.3 Resolution

- 546.1 nm wavelength is used
- FWHM of the peak is taken as the factor for the test.



1.3 Resolution

INSPECTION SHEET

Model: FP-6500
Serial No.: C089160822
Room temp.: 24.9 C
Humidity: 41 %
Operator: Ankit



Date: 30 / Mar / 2014

Resolution
Acceptance criterion

Overall Pass/Fail: Pass

(Ex) 1.10 nm
(Em) 1.10 nm
(Ex) 0.81 nm
(Em) 0.77 nm

Result(s)




1.4 Stray light

- Triangular cell with Rhodamine B dye used.
- In step 1 Rhodamine B is tested. [for emission]
- In step 2, only Diffusion plate is kept, no sample.
[for Excitation]



Exc.



EM.

Rhodamine B Solution

1.4 Stray light

INSPECTION SHEET



Model: FP-6500
Serial No.: C089160822
Room temp.: 25.0 C
Humidity: 41 %
Operator: Ankit

Date: 30 / Mar / 2014

Stray light
Acceptance criterion

Overall Pass/Fail: Pass
(Ex) Less than 3 %
(Em) Less than 1 %
(Ex) 0.30 %
(Em) 0.03 %

Result(s)




1.5 Minimum Detectable Amount

- Rectangular cell with Distilled water.
- We measure the time-course Raman peak at 398.4nm

Excitation at 350 nm
Emission at 397nm.

In wavenumbers,
we measure the Raman peak at
3380cm⁻¹ (O-H stretching)

S/n ratio is calculated after taking 10
measurements. The S/n ratio is the
criteria for the test.



1.5 Minimum Detectable Amount



INSPECTION SHEET

Model: FP-6500
Serial No.: C089160822
Room temp.: 25.0 C
Humidity: 41 %
Operator: Ankit

Date: 30 / Mar / 2014

Detection limit(Minimum detectable amount)	Overall Pass/Fail: Pass
Acceptance criterion	SN Ratio: More than 200(Ex 350 nm, Em 397 nm)
Result(s)	SN Ratio: 265.174 Noise: 0.87

The intensity of the peak at 3380cm^{-1} is recorded. Its stability over 15 sec for 10 cycles of measurement is analyzed.



1.6 Photometric Stability


- Rectangular cell with Distilled water.
- We measure the time-course Raman peak at 398.4nm

Excitation at 350 nm

Emission at 397.1 nm measured
over 1 hour.

In wavenumbers,
we measure the peak at 3380cm⁻¹.

Variation in the signal intensity is
recorded.



1.6 Photometric Stability



INSPECTION SHEET

Date: 30 / Mar / 2014

Model: FP-6500
Serial No.: C089160822
Room temp.: 24.6 C
Humidity: 41 %
Operator: Ankit

Photometric stability
Acceptance criterion
Result(s)

Overall Pass/Fail: Pass
2 %/hr(Ex 350 nm, Em 397.1 nm, 60 min)
0.00915567 %/hr



1. Correction


(by inbuilt software and given accessories)

- We can correct the Excitation intensity, Emission intensity, over the range 200-600nm using Rhodamine B solution.
For full spectral range correction we require Standard light source.

- Correction using Rhodamine B solution.

We measure the excitation of Rhodamine B over the range of 200-600nm;
emission wavelength is set at 640nm.

The sensitivity of the PMT is adjusted so that the peak value lies in between 500-800. Then we normalize at a particular wavelength.




- For our case:

We normalized at 400nm for both Excitation and Emission.

This was the first correction data produced.

And then we applied the correction data to compare with previous results.

Comparison after correction



I acknowledge Sudhakar Bhaiya for his
kind help and giving time on Sunday!